# PREVENTIVE MAINTENANCE CHECKLIST
# SIX (6) MONTH

**SERIAL NO.** _____ **DATE** _____

**MACHINE I.D. NO.** _____ **MACHINE HOURS** _____

**TECHNICIAN(S)** _____ **LABOR HOURS** _____

See the Referenced Section of the 2008T Preventive Maintenance Procedures (part number 508033 Rev. I) for detailed instructions to perform the following:

| Reference Section | Procedure | Value | Completed By: |
|---|---|---|---|
| 2.1 | **Clean Filters**.....(Replace If Necessary) | | |
| 2.2 | **Inspect Pre-UF Pump filter**.....(Replace If Necessary) | | |
| 2.3 | **Inspect UF Pump Check Valves**.....(Replace If Necessary) | | |
| 2.4 | **Check High Voltage AC Connections**.....(<0.2 ohms) | | |
| 2.5 | **Calibrate UF Pump**.....(Range: 24 ± 0.1cc at 24 UF Pump strokes) | | |
| 2.6 | **Check Conductivity Calibration**.....(Range: Δ ± 0.1mS/cm) | Machine Display: _____ | External Meter: _____ |
| 2.7 | **Check Temperature Calibration**.....(Range: 37 ± 0.3°C) | | |
| | (Range: 39 ± 0.3°C) | | |
| 2.8 | **Check Alarm Operation and Pressure Holding Test**..... | | |
| 2.9 | **Verify pH**.....(Range 6.9 to 7.6) | | |
| 2.10 | **Check Power Failure Alarm**..... | | |
| 2.11 | **Final Checks**..... | | |

**NOTES:**

_____

_____

_____

_____

_____

_____

_____

_____

Print Form