



**FRESENIUS  
MEDICAL CARE**

## Fresenius Medical Care North America Service Bulletin

Equipment: 2008T Hemodialysis Machines  
Bulletin: 12-FRT-006 Rev B  
Subject: CDX & Software Update Test Kit Upgrade (Update)

### 1.0 PURPOSE

To inform all Field Service and Technical Support personnel of how to upgrade the 2008T CDX & Software Update Test Kit with the latest Linux Operating (OS) system and instructions.

### 2.0 SCOPE

The 2008T CDX & Software Update Test Kit (P/N 190755) can be upgraded without the expense of a new router by using the 2008T CDX Test Kit Upgrade (P/N 190915):

#### **2008T CDX & Software Update Test Kit (P/N 190755\*) includes:**

- 190756 2008T USB CDX Test Script 3.0 Flash Drive
- 365075-03 Wireless Router
- 190731 CDX Compact Flash Card with OS
- 191141 CDX eSATA Solid State Drive with Linux OS
- 508419 2008T CDX & Software Update Test Instructions

#### **2008T CDX Test Kit Upgrade (P/N 190915\*) includes**

- 190731 CDX Compact Flash Card with Latest Linux OS
- 191141 CDX eSATA Solid State Drive with Latest Linux OS
- 508419 2008T CDX & Software Update Test Instructions

**\*Note:** The 2008T CDX & Software Update Test Kit (P/N 190755) is automatically updated to the latest CDX Compact Flash Card with OS. When ordering new test kit, a 2008T CDX Test Kit Upgrade (P/N 190915) is not needed.

If there are any questions regarding this bulletin, contact Fresenius Medical Care Technical Support at 800-227-2572.

© 2012 - 2018, Fresenius Medical Care North America, All rights reserved.